





# NABL

Department of Science & Technology, India

## SCOPE OF ACCREDITATION

Laboratory	Lawkim Limited, Mulund	Issue Date	03.12.2006
Field	Thermal Calibration	Valid Until	27.03.2008
Certificate Number	C-0065	Page	1 of 1
Last Amended on	-		

Quantity Measured / Instrument	Range	* Best Measurement Capability ( $\pm$ )	Remarks
<b><u>AT LABORATORY</u></b>			
1. RTD SENSOR WITH AND WITHOUT INDICATOR	-25°C – 35°C	0.1°C	Using PRT-PT-100(24460/6), 7½ DMM (40202) & ISOTECH Dry Block Furnace-21285/2
	35°C – 300°C	0.1°C	Using PRT-PT-100(24460/6), 7½ DMM (40202) & ISOTECH Dry Block Furnace-24460/4
	300°C – 600°C	0.1°C	Using PRT-PT-100(24460/6), 7½ DMM (40202) & ISOTECH Dry Block Furnace-212285-1-2
2. T/C SENSOR WITH INDICATOR	600°C – 1000°C 1000°C – 1200°C	1.8°C 4.2°C	Using S-Type Thermocouple, Eurotherm Temp. Indicator (LQE022) & Dry Block Furnace 212285-1 or 24460/2
<b><u>AT SITE</u></b>			
3. RTD SENSOR WITH INDICATOR	-25°C – 35°C	0.6°C	Using PT-100 (RSPL-02), Eurotherm Temp. Indicator (LQE-021) & ISOTECH Dry Block Furnace 21228/2
	35°C – 300°C	0.8°C	Using PT-100 (RSPL-02), Eurotherm Temp. Indicator & ISOTECH Dry Block Furnace 24460/5
	300°C – 600°C	0.9°C	Using S-Type T/C (MTC-0650), Eurotherm Temp. Indicator & ISOTECH Dry Block Furnace 24460-1
4. T/C SENSOR WITH INDICATOR	600°C – 1000°C	1.8°C	Using S-Type T/C (MTC-0650), Eurotherm Temp. Indicator & ISOTECH Dry Block Furnace 24460-1

\* Measurement Capability is expressed as an uncertainty ( $\pm$ ) at a confidence probability of 95 %

*Rubana Saloom*  
Convenor